

Multi-Metals Continuous Water Analyzer Based on ED-XRF: Applications to Power Plant ELG Rule Compliance

**Krag Petterson, Troy Pittenger, and Brian Edge
Cooper Environmental**

Presentation Outline

- Summary of Current State of Effluent Limitation Guidelines for Steam EGU
- Advantages of Real Time Monitoring
- Operation and Capabilities of Xact 920
- Laboratory Testing on Xact
- Field Testing monitoring Se in Bioreactor Effluent

Effluent Limit Guideline (ELG) Rule Summary

- In November of 2019, EPA proposed changes to 2015 Effluent Limitations Guidelines and Standards for the Steam Electric Power Generating Point Sources
- These proposed changes affected the rules for two types of power plant waste streams
 - Flue Gas Desulfurization (FGD) Wastewater
 - Bottom Ash (BA) Transport Water
- For FGD wastewater the rule creates essentially five subcategories of sources – each with its own limits
 - Normal or typical FGD treatment
 - High flow FGD facilities
 - Low utilization boilers
 - Boilers retiring by 2028
 - Voluntary incentives program (VIP)

ELG Rule Summary – Limits for Each Category

Subcategory	Pollutant	Long-Term Ave	Daily Max	Monthly Ave
Typical Facilities	Arsenic (µg/L) (ppb)	5.1	18	9
	Mercury (ng/L) (ppt)	13.5	85	31
	Nitrate/Nitrite as N (mg/L) (ppm)	2.6	4.6	3.2
	Selenium (µg/L) (ppb)	16.6	76	31
Voluntary Incentives Program for FGD Wastewater	Arsenic (µg/L) (ppb)	5.0	5	-
	Mercury (ng/L) (ppt)	5.1	21	9
	Nitrate/Nitrite as N (mg/L) (ppm)	0.4	1.1	0.6
	Selenium (µg/L) (ppb)	5.0	21	11
	Bromide (mg/L) (ppm)	0.16	0.6	0.3
	TDS (mg/L) (ppm)	88	351	156
Low Utilization and High Flow	Arsenic (µg/L) (ppb)	5.98	11	8
	Mercury (ng/L) (ppt)	159	788	356

ELG Rule Summary

- Currently compliance would be required as “as soon as possible” once the rule is promulgated
- Most plants will need to use bioreactor technology to achieve Se limits
- Daily measurements of Se, As and Hg are required
- Plants not removed from service or in the Voluntary Incentives Program (VIP) must meet the limits by December 2025
- VIP plants must meet by December 2028

Why Measure Pollutants in Real Time

- Disadvantages of Manual Measurements
 - Slow – results from commercial labs take days or even weeks
 - Expensive – on per sample basis results are expensive especially for quick turns
 - Can't get immediate feedback to know if adjustments to control strategies are working
 - Don't know until you may have exceeded a daily or monthly limit until its too late to do anything about it

Why Measure Pollutants in Real Time


- Advantages of Real Time Analysis results
 - Can determine if an ELG limit exceedance will take place before it does
 - Allows plant operators to adjust controls to prevent an ELG limit exceedance
 - Allows for immediate feedback so that operators can fine tune controls to make them operate more efficiently – saving the power plant money
 - Minimize human error in the analysis process
 - Could do inlet and outlet measurements from bioreactor to determine treatment efficiency

Xact 920 Continuous Metals In Water Analyzer



- Xact 920 can measure up to 65 elements simultaneously including **Se**, **As**, and **Br**
- Water is spray dried and sampled onto filter tape
- The resulting filter tape deposit is analyzed by X-ray Fluorescence (XRF)
- Builds on two technologies developed and commercialized by Cooper Environmental
 - Ambient Air XRF Analyzer (over 150 in field)
 - Quantitative Aerosol Generator (developed for calibration of PM CEMS for the power industry with EPRI Support)

XRF Theory



- Incoming X-rays eject an inner shell electron
- Electrons from higher shells fill the vacancy
- This process releases energy in the form of fluorescing X-rays
- Energy is characteristic of each element
- Intensity or brightness is related to the mass of each element

Strengths of XRF

- XRF utilizes inner shell electron transitions so the response is not dependent on what is chemically bound to the element
- Can measure a wide range of elements simultaneously
- XRF is non-destructive – so samples can be reanalyzed later
- XRF is very stable – calibrations can last for years
- XRF response is linear over a wide concentration range (over 5 orders of magnitude) – this means no additional standards required depending on concentration range

Measurable Elements

 Elements measurable by the Xact 920

 Elements with Limits under the ELG Rule

 Other Elements of Potential Interest in the Power Industry


1 H																	2 He
3 Li	4 Be											5 B	6 C	7 N	8 O	9 F	10 Ne
11 Na	12 Mg											13 Al	14 Si	15 P	16 S	17 Cl	18 Ar
19 K	20 Ca	21 Sc	22 Ti	23 V	24 Cr	25 Mn	26 Fe	27 Co	28 Ni	29 Cu	30 Zn	31 Ga	32 Ge	33 As	34 Se	35 Br	36 Kr
37 Rb	38 Sr	39 Y	40 Zr	41 Nb	42 Mo	43 Tc	44 Ru	45 Rh	46 Pd	47 Ag	48 Cd	49 In	50 Sn	51 Sb	52 Te	53 I	54 Xe
55 Cs	56 Ba	55-71 *	72 Hf	73 Ta	74 W	75 Re	76 Os	77 Ir	78 Pt	79 Au	80 Hg	81 Tl	82 Pb	83 Bi	84 Po	85 At	86 Rn
87 Fr	88 Ra	89-103 †	104 Rf	105 Db	106 Sg	107 Bh	108 Hs	109 Mt	110 Ds	111 Rg	112 Cn	113 Uut	114 Fl	115 Uup	116 Lv	117 Uus	118 Uuo

*Lanthanides
(rare earth metals)


†Actinides

57 La	58 Ce	59 Pr	60 Nd	61 Pm	62 Sm	63 Eu	64 Gd	65 Tb	66 Dy	67 Ho	68 Er	69 Tm	70 Yb	71 Lu
89 Ac	90 Th	91 Pa	92 U	93 Np	94 Pu	95 Am	96 Cm	97 Bk	98 Cf	99 Es	100 Fm	101 Md	102 No	103 Lr


General Operation Schematic - Xact 920



Instrument Systems




XRF Sampling and Analysis



Xact 920 – Quality Assurance


- XRF Portion Calibrated with Traceable to NIST Thin Film Standards
- Automatic XRF spectrometer stability check with every sample
- Daily upscale check of XRF
- Stability – XRF calibration frequency about once/year – sometimes years between calibrations



Laboratory Testing




Laboratory Results – Low Concentration Drift



Metal	Zero Drift	Average Measured Value (ng/g)	Standard Deviation (ng/g)	N	% RSD	Actual Value (ng/g)	RPE
Ni	0.3%	525.6	3.6	75	0.7%	501.8	-5%
Cu	-0.5%	463.3	4.0		0.9%	508.1	9%
Zn	-0.2%	483.4	3.5		0.7%	501.8	4%
As	0.1%	506.5	3.4		0.7%	505.7	0%
Cd	0.8%	550.7	9.3		1.7%	497.7	-11%
Pb	0.3%	521.7	4.7		0.9%	509.4	-2%

- Xact 920 shows excellent stability over the course of the two day test
- The relative standard deviation for most elements was less than 1%

Laboratory Results – Span Drift





**30 MINUTE
MODE**

Metal	Span Drift	Average Measured Value (ng/g)	Standard Deviation (ng/g)	N	% RSD	Actual Value (ng/g)	RPE
Ni	-1.7%	8498.5	47.6	84	0.6%	7930	-7.2%
Cu	-1.9%	7479.4	39.7		0.5%	7994	6.4%
Zn	1.4%	7760.4	40.7		0.5%	7894	1.7%
As	-1.5%	8117.8	48.8		0.6%	7957	-2.0%
Cd	-2.6%	8320.3	117.9		1.4%	7831	-6.3%
Pb	1.4%	8397.9	50.7		0.6%	8015	-4.8%

Laboratory Results – Linearity

Ni

Zn





NO Change In Calibration Over Solution Concentration Range

Xact Shows Linearity Over a Range


Spanning 5 Orders of Magnitude

Cu

As




Low Concentration Level Accuracy Testing



Demonstrated Measurement Accuracy down to 1 ppb


- Testing done as a factory acceptance test for Nuclear Power Plant Application
- Compared the Xact 920 response to a known solution concentration at very low concentration levels
- The slope of the best fit line is very nearly 1.0 indicating good agreement between the Xact reported concentration and the known solution concentration

Field Testing



Xact 920 Field Evaluation

- Xact 920 sampling bioreactor effluent at a large (multi boiler) power plant
- Instrument installed January 6, 2020 – operational within 4 hours
- Sampling from bioreactor effluent holding tank



Plant FGD Wastewater Control Process Schematic

Xact 920 Field Evaluation Study

- Xact operated continuously from January 9th to March 13th
- This presentation includes data from January 9 to March 13th
- Uptime of 97.5% in the data time frame
- Bioreactor operation started approximately 2/19/2020
- Xact results compared to laboratory (Brooks Lab, Bothell Washington)
- Two different types of samples
 - Analysis of the same grab sample – Xact analyzes the same sample as shipped to lab
 - A grab sample acquired while the Xact was operating continuously
- Three grab sample campaigns
 - 1/7/2020 to 1/9/2020
 - 1/24/2020 to 1/25/2020
 - 2/20/2020 to 2/24/2020


Lab Comparison Results

Xact 920 - Percent Error Comparison between Xact and ICP-MS								
Identification		Grab ID and Sample Prep			Xact 920	Lab Data	Comparison	
Lab Analysis Group	Lab ID	Date	Time	Location	N	Avg Se Conc. (ppb)	Se Conc. (ppb)	Xact % Error
A	010720a	1/7/20	13:00	Clarifier Overflow	6	26.99	31.18	-13.4%
	010920b	1/9/20	9:30	Clarifier Overflow	66	21.60	22.5	-4.0%
	010920c	1/9/20	9:30	Clarifier Overflow	8	126.0	119	5.9%
	010920d	1/9/20	9:30	Clarifier Overflow	5	106.0	127	-16.5%
	010920e	1/9/20	14:00	SeHAWK BioEff 2A	5	61.53	59.3	3.8%
	011020a	1/9/20	14:00	SeHAWK BioEff 2A	6	79.28	80.3	-1.3%
	010920f	1/9/20	14:00	SeHAWK BioEff 2B	5	65.32	73.4	-11.0%
B	012720c	1/24/20	10:38	SeHAWK BioEff Tank	1	46.04	51.2	-10.1%
	012620a	1/24/20	12:25	SeHAWK BioEff Tank	7	52.94	48.9	8.3%
	012620b	1/24/20	12:25	SeHAWK BioEff Tank	7	152.2	153	-0.5%
	012520a	1/24/20	16:25	SeHAWK BioEff Tank	7	51.84	46.5	11.5%
	012520b	1/24/20	16:25	SeHAWK BioEff Tank	6	141.9	127	11.8%
	012720b	1/24/20	17:25	SeHAWK BioEff Tank	1	48.74	51.1	-4.6%
	012720a	1/25/20	15:05	SeHAWK BioEff Tank	18	99.47	101	-1.5%
C	200124.1600	1/24/20	16:00	Clarifier Overflow	22	95.29	97.7	-2.5%
	200220.1235	2/20/20	12:35	SeHAWK BioEff Tank	1	128.2	129	-0.6%
	200220.1505	2/20/20	15:05	SeHAWK BioEff Tank	1	129.2	129	0.2%
	200221.0805	2/21/20	8:05	SeHAWK BioEff Tank	1	107.8	107	0.8%
	200221.1240	2/21/20	12:40	SeHAWK BioEff Tank	1	107.0	107	0.0%
	200221.1605	2/21/20	16:05	SeHAWK BioEff Tank	1	79.25	78.4	1.1%
	200222.1905	2/22/20	19:05	SeHAWK BioEff Tank	1	76.07	80.7	-5.7%
	200223.1230	2/23/20	12:30	SeHAWK BioEff Tank	1	52.01	53.2	-2.2%
200224.1005	2/24/20	10:05	SeHAWK BioEff Tank	1	44.25	44.3	-0.1%	
Total Number of Lab Comparison Samples							23	
Total Average Percent Difference with the Lab							-1.3%	
Standard Deviation of Percent Difference							7.1%	

- Tested concentrations range from about 20 ppb to about 150 ppb (includes spike samples)

- Average Percent Error is -1.3%

Bioreactor Effluent Parity Plot with Laboratory



NO data has been excluded

- The slope shows excellent agreement with laboratory ICP-MS
- The R^2 for this also shows excellent precision between the Xact and laboratory analysis

Spike Recovery

Spike Recovery Summary Xact 920 and ICP-MS										
Identification		Grab ID and Sample Prep				Spike			Xact 920	ICP-MS
Series	Lab	Date	Time	Location	Filter (um)	Se Spike (ppb)	Spike Type	N	Spike Recovery (%)	Spike Recovery (%)
A	010920c	1/9/20	9:30	Clarifier Overflow	250	100.4	SeVI	8	105.8%	97.8%
	010920d	1/9/20	9:30	Clarifier Overflow	250	102.7	SeIV	5	83.9%	103.8%
	011020a	1/9/20	14:00	SeHAWK BioEff 2A		21.40	SeVI	6	83.3%	98.5%
B	012620b	1/24/20	12:25	SeHAWK BioEff Tank	250	109.7	SeIV	7	92.5%	97.0%
	012520b	1/24/20	16:25	SeHAWK BioEff Tank	250	85.53	SeVI	6	106.9%	95.5%

Notes:

- EPA Method 200.8 matrix spike recovery range is 70 to 130%
-
- Spike Recovery = $(\text{SpikeResult} - \text{BaselineResult}) / (\text{SpikeConc} * \text{DilutionFactor})$

- US EPA recommended range for spike recovery is between 70% and 130%
- Both the Xact and ICP-MS percent recoveries fall within this range
- Xact and ICP-MS spike recovery is within range for both Se (VI) and Se (IV)

Detection Limit Determination

Xact 920 - Minimum Detection Limit Calculations for Clarifier Overflow Sample Matrix											
Meta		Measured Concentration (ppb Se)							DL Calculation (ppb Se)		
ID	Date	Rep 1	2	3	4	5	6	7	StDev	99% DL	Avg. 99% DL
1	1/15/20	22.89	21.44	22.18	21.51	22.66	21.06	21.62	0.68	2.14	3.12
2	1/16/20	22.71	22.66	21.60	21.24	20.33	22.97	21.63	0.95	3.00	
3	1/17/20	21.10	20.91	22.16	19.89	19.97	22.49	22.12	1.06	3.34	
4	1/20/20	23.54	23.02	21.17	23.19	21.68	22.01	22.19	0.87	2.73	
5	1/21/20	20.37	22.41	20.18	20.38	21.42	19.89	17.88	1.40	4.39	

Notes:

1. Single-tailed student t-value for 99% confidence level, for 7 replicates equals 3.143
2. The replicate measurements are the first seven replicates of each day

**Clarifier
Overflow
Detection
Limit
(ppb)**

Xact 920 - Minimum Detection Limit Calculations for Bioreactor Effluent Sample Matrix											
Meta		Measured Concentration (ppb Se)							DL Calculation (ppb Se)		
ID	Date	Rep 1	2	3	4	5	6	7	StDev	99% DL	Avg. 99% DL
1	3/14/20	4.20	4.98	3.42	5.90	4.32	3.71	5.15	0.87	2.73	2.55
2	3/15/20	5.86	6.05	7.82	6.22	6.77	6.71	5.40	0.79	2.47	
3	3/16/20	7.89	8.64	7.71	8.75	7.96	8.83	8.61	0.47	1.47	
4	3/17/20	6.61	7.54	9.44	8.73	8.67	9.91	8.74	1.12	3.52	


Notes:

1. Single-tailed student t-value for 99% confidence level, for 7 replicates equals 3.143
2. The replicate measurements are the first seven replicates of each day

**Bioreactor
Effluent
Detection
Limit
(ppb)**


- Detection limit determined based on your sample matrix –

Xact Operational Data Versus Laboratory




- Data is from short term operation of Bioreactor from 1/23 to 1/25
- During normal instrument operation the Xact data matches the lab data very well
- Xact data matches laboratory data during bioreactor Shutdown

Xact Operational Data Versus Laboratory




- Data is from operation of bioreactor from 2/20 to 2/25
- During normal instrument operation the Xact data matches the lab data very well
- Xact data matches laboratory data during bioreactor startup

Concentration During all Bioreactor Operation



- There are times when Se concentration data does not correlate with changes in bioreactor effluent operation
- Understanding these changes could lead to more efficient bioreactor operations

Other Elements measured by the Xact



- Note log scale
- Xact able to measure other elements besides Se

Conclusions

- Real time data can provide useful feedback for plants trying to comply with ELG limits
- Xact 920 provides data that is accurate, reliable and compares well with independent lab analysis
- Xact 920 is easy to install and operates very reliably

Questions?

Email: Kragp@cooperenvironmental.com